

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

Page 1 of 2

VERIFICATION OF CALIBRATION REPORT
Of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory -- Statewide Breath Alcohol Program

DataMaster cdm S/N 130296 ✓

Supervisor/Operator & Number Performing the Verification Procedure: John J. Waldron, #3392
Department and Date: Unalaska DPS,

(CONTINUED FROM FRONT PAGE)

C. TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130296 ✓
SEPTEMBER 16, 2009 ✓ *Bmb 10/2/09*
OPERATOR'S NAME: WALDRON/JOHN/J
OPERATOR'S NUMBER: 3392
SUBJECT'S LAST NAME: UUC
SUBJECT'S FIRST NAME/MI: UUC
O.I.L. #: 01234567
CERT AGENCY: UNK1
TEST REPORT: 09-00000
TEST TYPE: V
ALCO TARGET VALUE: .002 ✓
ALCO S/N: 78966 ✓
--- BREATH ANALYSIS ---
.002 ADJUSTED FOR 29.47 in ✓
ALCO TARGET .000 02:16
BLANK TEST .000 02:17
INTERNAL STANDARD VERIFIED 02:17
ALCO TU 29.47 in .903 ✓ 02:18
BLANK TEST .000 02:18
SUBJECT SAMPLE *Bmb 10/2/09* .000 02:19
BLANK TEST .000 02:20
ALCO TU 29.47 in .003 ✓ 02:20
BLANK TEST .000 02:21

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130296 ✓
SEPTEMBER 16, 2009 ✓
TIME 02:21
--- DIAGNOSTIC CHECK ---
COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 48c
BAROMETER: 29.47 in
FLOW DETECTOR: OKAY
PUMP
MOTOR SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY
PRINTER TEST
!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~■